Occupational hygiene monitoring of dust

Principle

The procedure utilizes sampling cartridges with pre-weighted filters. When air is sucked through the sampling cartridge using a calibrated precision air sampling pump, airborne dust is collected on the filter. During analysis the filter is weighed and the air concentration of dust calculated.





Contact: Nils Arne Jentoft

Analytical Services Manager

Phone: +47 90 87 34 70 E-mail: naj@dynea.com